UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/624,581	07/22/2003	Holger Soring	SOR 205	7516
7590 02/27/2007 Horst Kasper 13 Forest Drive Warren, NJ 07059			EXAMINER	
		KOHARS ART UNIT	KOHARSKI, C	HRISTOPHER
			ART UNIT	PAPER NUMBER
			3763	
Г 				
SHORTENED STATUTORY PERIOD OF RESPONSE		MAIL DATE	DELIVERY MODE	
31 DAYS		02/27/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

1		W.				
	Application No.	Applicant(s)				
Office Action Summans	10/624,581	SORING ET AL.				
Office Action Summary	Examiner	Art.Unit				
	Christopher D. Koharski	3763				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	correspondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tim will apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).				
Status						
1) Responsive to communication(s) filed on 24 No	ovember 2006.					
2a) This action is FINAL . 2b) ☑ This						
3) Since this application is in condition for allowar	nce except for formal matters, pro	secution as to the merits is				
closed in accordance with the practice under E	x parte Quayle, 1935 C.D. 11, 45	53 O.G. 213.				
Disposition of Claims						
4) Claim(s) 1-32 is/are pending in the application.						
4a) Of the above claim(s) is/are withdrav	4a) Of the above claim(s) is/are withdrawn from consideration.					
5) Claim(s) is/are allowed.						
6) Claim(s) is/are rejected.						
7) Claim(s) is/are objected to.	Jantian saguisamant					
8) Claim(s) <u>1-32</u> are subject to restriction and/or e	election requirement.					
Application Papers						
9)☐ The specification is objected to by the Examine	r.	·				
10)☐ The drawing(s) filed on is/are: a)☐ accepted or b)☐ objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the Ex	•					
	ammer. Note the attached Office	ACTION OF IGHT F 10-132.				
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of:						
	 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 					
3. Copies of the certified copies of the prior						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)						
1) Notice of References Cited (PTO-892)	4) Interview Summary					
Notice of Draftsperson's Patent Drawing Review (PTO-948) Information Disclosure Statement(s) (PTO/SB/08)	Paper No(s)/Mail D 5) Notice of Informal F					
Paper No(s)/Mail Date	6) Other:					

Application/Control Number: 10/624,581

Art Unit: 3763

DETAILED ACTION

Response to Amendment

Examiner acknowledges the reply filed 11/24/2006, due to a file error claims 30-32 were omitted from prior action and are included in the new groupings as specified below. Additionally, the last response from the Applicant's representative was non-compliant, although the restriction was traversed; Applicant must still chose an inventive group. Therefore the following updated restriction is presented below.

Election/Restrictions

Restriction to one of the following inventions is required under 35 U.S.C. 121:

- Claims 1-17 and 27-32, drawn to medical treatment apparatus with an ultrasound generator hand piece with sonotrode and a suction element, classified in class 604, subclass 20.
- II. Claims 18-26, drawn to medical treatment apparatus with an ultrasound generator hand piece with sonotrode and flushing line and container, classified in class 604, subclass 22.

Inventions I and II related as subcombinations disclosed as usable together in a single combination. The subcombinations are distinct if they do not overlap in scope and are not obvious variants, and if it is shown that at least one subcombination is separately usable. In the instant case, subcombinations of Group I and II have separate utility such as the device (II) can be used to administer medical treatment along with a medical solution or the device (I) can

be used to administer medical treatment and remove excess fluids or gases from the treatment site via the suction means, these activities have separate utility and have separate status in the art. See MPEP § 806.05(d).

The examiner has required restriction between subcombinations usable together. Where applicant elects a subcombination and claims thereto are subsequently found allowable, any claim(s) depending from or otherwise requiring all the limitations of the allowable subcombination will be examined for patentability in accordance with 37 CFR 1.104. See MPEP § 821.04(a). Applicant is advised that if any claim presented in a continuation or divisional application is anticipated by, or includes all the limitations of, a claim that is allowable in the present application, such claim may be subject to provisional statutory and/or nonstatutory double patenting rejections over the claims of the instant application.

Because these inventions are independent or distinct for the reasons given above and there would be a serious burden on the examiner if restriction is not required because the inventions have acquired a separate status in the art in view of their different classification, restriction for examination purposes as indicated is proper.

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

Application/Control Number: 10/624,581

Art Unit: 3763

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.

Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Koharski whose telephone number is 571-272-7230. The examiner can normally be reached on 7:30am to 4:00pm EST.

Art Unit: 3763

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Date:

14 004 Christopher D. Koharski AU 3763